

# Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

STERILITY LOT NO	PRODUCT RELEASE DATE	STERILITY CERT LOT
CS20221212-US	20230323	2300454-1

DOSE INFO	MIN Specified	MAX Specified	MIN Actual	MAX Actual
(kYg)	30.0	50.0	32.2	42.1

MFG LOT NUMBER(S)	M2300285							

PRODUCT ITEM NO	SIZE							
CTL428CS	XL							

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [www.lakeland.com](http://www.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).



M. LEE MORGAN, Ph.D.

Global Director, Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel



[Info@lakeland.com](mailto:Info@lakeland.com)  
[www.lakeland.com](http://www.lakeland.com)  
800-645-9291

Certificate of Sterility