

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

STERILITY LOT NO	PRODUCT RELEASE DATE	STERILITY CERT LOT
CS20230118-US	20230323-0409	2300454-3 2044576-2

DOSE INFO	MIN Specified	MAX Specified	MIN Actual	MAX Actual
(kYg)	30.0	50.0	31.3 30.0	42.0 46.7

MFG LOT NUMBER(S)	M2300289	M2300286	M2300358	M2300359	M2300360	M2300361	M2300362	M2300363
	M2300364							

PRODUCT ITEM NO	SIZE						
CTL428CS	XL	4X					
CTL417CS	SM						
CTL191CS	XL	2X					
CTL428CS	2X	5X					
CTL903CSP	SMMD	LGXL					

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.



M. LEE MORGAN, Ph.D.

Global Director, Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



Info@lakeland.com
www.lakeland.com
800-645-9291

Certificate of Sterility