

# Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

STERILITY LOT NO	PRODUCT RELEASE DATE	STERILITY CERT LOT
CS20230221-US	20230420	2300577

DOSE INFO	MIN Specified	MAX Specified	MIN Actual	MAX Actual
(kYg)	30.0	50.0	31.5	42.6

MFG LOT NUMBER(S)	M2300405	M2300406	M2300407	M2300408	M2300409	M2300410	M2300411	

PRODUCT ITEM NO	SIZE						
CTL191CS	2X						
CTL417CS	XL	2X					
CTL428CS	XL	2X	3X	4X			

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [www.lakeland.com](http://www.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

  
M. LEE MORGAN, Ph.D.

Global Director, Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel



[Info@lakeland.com](mailto:Info@lakeland.com)  
[www.lakeland.com](http://www.lakeland.com)  
800-645-9291

Certificate of Sterility