

# Certificate of Sterility

## Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

STERILITY LOT NO	PRODUCT RELEASE DATE	STERILITY CERT LOT
CS20230330-US	20230504	2300742

DOSE INFO	MIN Specified	MAX Specified	MIN Actual	MAX Actual
(kGy)	27.5	42.0	28.3	36.9

MFG LOT NUMBER(S)	M2300501	M2300502	M2300503	M2300504	M2300505	M2300506	M2300507	M2300508

PRODUCT ITEM NO	SIZE						
CTL417CS	MD	LG	4X				
CTL428CS	SM	XL	2X	3X	4X		

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [www.lakeland.com](http://www.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

  
M. LEE MORGAN, Ph.D.

Global Director, Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel



[Info@lakeland.com](mailto:Info@lakeland.com)  
[www.lakeland.com](http://www.lakeland.com)  
800-645-9291