

Certificate of Sterility

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

STERILITY LOT NO	PRODUCT RELEASE DATE	STERILITY CERT LOT
CS20230422-US	20230618	2300973-1 AND 2300973-2

DOSE INFO	MIN Specified	MAX Specified	MIN Actual	MAX Actual
(kGy)	27.5	42.0	27.5-27.8	42.0-38.6

MFG LOT NUMBER(S)	M2300682	M2300683	M2300684	M2300685	M2300686	M2300687	M2300688	M2300689
	M2300664	M2300665	M2300666	M2300667	M2300668	M2300669	M2300670	M2300671
	M2300672							

PRODUCT ITEM NO	SIZE						
CTL417CS	XL						
CTL428CS	LG	XL	2X	3X	4X	5X	
CTL903CSP	LGXL						
CTL191CS	LG	XL	2X				
CTL414CS	MD	LGXL	XL				
CTL417CS	MD	LG	2X				

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10⁻⁶. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.



M. LEE MORGAN, Ph.D.

Global Director, Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



Info@lakeland.com
www.lakeland.com
800-645-9291